

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 09 2009

JUL 20 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130130</u>
Name <u>JEREMY COOKLIN</u>		ID# <u>6715</u> Date <u>07.02.09</u>
A	Agency <u>VALDEZ POLICE DEPARTMENT</u>	Phone # <u>835-4560</u>
Instrument Location <u>VALDEZ JAIL</u>		
Alco S/N <u>X124630</u>		Target Value <u>.078</u> High Pressure <u>1250</u>
B	Alco Test Value Average <u>.079</u> ✓ 1 st Alco <u>.079</u> ✓ 2 nd Alco	
Signature <u>[Signature]</u> <u>271</u> <u>083</u> <u>7/24/09</u>		
(OVER)		

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


(Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 24th day of July 2009.

Carolyn M. Noland
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

JULY 02, 2009

OPERATOR'S NAME:

COMPLING

OPERATOR'S NUMBER: 5715

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI:

UOC

O.L. #1 UOC

LEFT BLENDAIR 10021

ALOE REPORT: 100

TEST TUBE 2

BREATH ANALYSIS

.079 ADJUSTED FOR 30.64 in		
ALCO TARGET	.079	01:53
BLANK TEST	.000	01:54
INTERNAL STANDARD	VERIFIED	01:54
ALCO TV 30.64 in	.079	01:55
BLANK TEST	.000	01:55
SUBJECT SAMPLE	.000	01:56
BLANK TEST	.000	01:57
ALCO TV 30.64 in	.079	01:57
BLANK TEST	.000	01:58

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

JULY 02, 2009

TIME 01:51

DIAGNOSTIC CHECK

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 30.64

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

ALCO TV 30.64 in .079
BLANK TEST .000
INTERNAL STANDARD VERIFIED
ALOE REPORT: 100
TEST TUBE 2